

Analytical Report Nr.

AR-20-YL-007362-01

Sample code Nr.

560-2020-00007819

Date

06/11/2020

ANALYTICAL REPORT**Client Information**

Fergotex – Fábrica de Malhas Têxteis, Lda.
Av. Conde de Arnoso, 1831
Vila Nova de Famalicão PORTUGAL
+351 252910200
carlospereira@fergotex.pt
For the attention of Mr Carlos Pereira

Sample Information

Order Code: EUAA70-00008926
Reception Date: 26-Oct-2020
Analysis Starting Date: 26-Oct-2020
Analysis Ending Date: 3-Nov-2020
Sample code Nr. 560-2020-00007819
Sample described as: Face mask

Requirements and decision rule

Customer requirements: CWA 17553:2020 Community face coverings requirements- Level 70%
Decision Rule: Shared risk - Simple acceptance.

Information provided by the customer*

Client Reference: FRGMSK -01
Sample Description: Community mask CWA 17553
Purchase Order Number:

Batch Not provided

Analytical Report Nr.

AR-20-YL-007362-01

Sample code Nr.

560-2020-00007819

Date

06/11/2020

SAMPLE PICTURE

Analytical Report Nr.

AR-20-YL-007362-01

Sample code Nr.

560-2020-00007819

Date

06/11/2020

CONCLUSION:

TEST PROPERTY	PASS	FAIL	REMARKS
• Bacterial Filtration Efficiency (BFE) EN 14683:2019+AC:2019 Annex B			
A-Mask	X		
Breathability (Differential Pressure) EN 14683:2019+AC:2019 Annex C			
A-Mask	X		
Resistance against penetration by synthetic blood ISO 22609:2004			
A-Mask			REFER RESULT

Remark: Test has been performed as per application request

Analytical Report Nr.

AR-20-YL-007362-01

Sample code Nr.

560-2020-00007819

Date

06/11/2020

COMPONENT LIST:

COMPONENT ID	COMPONENT NAME	MATERIAL DESCRIPTION	COLOR	REMARKS
CUST 01	A-Mask	Face mask	Blue	---

Analytical Report Nr.

AR-20-YL-007362-01

Sample code Nr.

560-2020-00007819

Date

06/11/2020

MASKS TESTING	CAS No.	RESULTS	UNC.	LOQ	GUIDELINES
---------------	---------	---------	------	-----	------------

Analyses on:A-Mask

Resistance against penetration by synthetic blood

Analysis Ending Date: 03/11/2020

ISO 22609:2004

Result: 3 of 32 samples pass at 16 kPa

-

Complete test data reported at Annex.

AQL information according to ISO 22609:2004:

A single sampling plan providing an AQL of 4,0 % requires 32 test specimens.

An AQL of 4.0 % is met for a single sampling plan when 29 or more specimens show pass results.

AQL= Acceptable Quality Limit.

Breathability (Differential Pressure)

Analysis Ending Date: 03/11/2020

EN 14683:2019+AC:2019 Annex C

Differential pressure

18.6 Pa/cm² (± 1.0) Pa/cm²

-

<60 Pa/cm²

✓ Pass

Complete test data reported at Annex.

• Bacterial Filtration Efficiency (BFE)

Analysis Ending Date: 03/11/2020

EN 14683:2019+AC:2019 Annex B

Bacterial Filtration Efficiency (BFE)

83.52 %

-

≥ 70 %

✓ Pass

Complete test report attached as annex.

Test covered by ACCREDIA accreditation scope n° 1827 L

Analytical Report Nr.

AR-20-YL-007362-01

Sample code Nr.

560-2020-00007819

Date

06/11/2020

Signed for and on behalf of Eurofins Textile Testing Spain:
Eurofins Textile Testing Spain, S.L.U.
C/ Germán Bernácer 4, Elche (Alicante)
030078009

Report electronically validated by

Luis Miralles Esteve

Quality Manager

EXPLANATORY NOTE

- ◆ Test not covered by ENAC accreditation scope
- Test is subcontracted within Eurofins group and is accredited
- Test is subcontracted within Eurofins group and is not accredited
- Test is subcontracted outside Eurofins group and is accredited
- Test is subcontracted outside Eurofins group and is not accredited

N/A = Not Applicable

Eurofins Textile Testing Spain S.L.U is not responsible of the information supplied by the costumer and reported as section "Information provided by the costumer".

Eurofins General Sales Terms and Conditions Applied.

Results obtained refer only to samples, products or material received in Laboratory, as described in section "Sample information" and tested in conditions shown in present report.

Test uncertainties not reported are at customer disposal, for those tests in which it is possible to evaluate the test uncertainty.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution provides a level of confidence of approximately 95%.

Reproduction of this document is only valid if it is done completely and under the written permission. Any printed version of this document will be considered as a copy.

If you happen to have any comments, please do it by sending email to textile_spain@eurofins.com and referring to this report number.

End Of Report

Eurofins Textile Testing Spain, S.L.U.

Calle Germán Bernácer, 4

03203 Elche

SPAIN

Phone+34 966 299 638**www.eurofins.com/tex**

ENAC is signatory of EA and ILAC Multilateral Agreement for testing

Activities not covered by ENAC accreditation are marked with ◆ ○ ● □ ■

METHOD FOR DETERMINATION OF BREATHABILITY (DIFFERENTIAL PRESSURE)

Test Method: EN 14683: 2019+AC: 2019 Annex C

Number of test specimens: 5


Number of test per specimen: 5

Sample area tested: Circular, diameter 2,5 cm

Tested area of the test sample: 4,9 cm²

Flow rate during testing: 8±0,25 l/min

General location of measurement areas: Representative of the overall surface.



Results

Specimen	Units (Pa)						ΔP (Pa/cm ²)
	Position 1	Position 2	Position 3	Position 4	Position 5	Mean value (Pa)	
1	88	85	90	91	85	88	17,9
2	95	88	91	89	88	90	18,4
3	91	87	98	92	95	93	18,9
4	92	86	95	86	94	91	18,5
5	94	93	94	96	94	94	19,2
Mean Value							18,6
Uncertainty							± 1,0

Observation:

For thick and rigid masks the test method may not be suitable as a proper seal cannot be maintained in the sample holder.

DETERMINATION RESISTANCE AGAINST PENETRATION BY SYNTETIC BLOOD

Test Method: ISO 22609:2004; Targeting-plate test method

Number of test specimens: 32

Sample size: Circular, diameter 5,58 cm

Sample area tested: 24,5 cm²

Pressure: 16 kPa (120,0 mm Hg)

Stream velocity of synthetic blood: 550±10 cm/s

Distance of the face mask target area surface from the tip of the cannula: 30,5 cm

Angle of the pneumatic valve with respect to the face mask target area: 90°

Technique used to enhance visual detection of synthetic blood: Hydrophilic cotton

Conditioning: At least 4 hours. T^a between 16,7°C and 26 °C. HR% between 82,8 and 88 % Hr

Environmental test conditions 23,0 °C; 84,2 % Hr

Pre-treatment: None

Results		
Specimen	Pass	Fail
1		X
2		X
3		X
4		X
5		X
6		X
7		X
8	x	
9		X
10		X
11		X
12		X
13		X
14		X
15		X
16		X
17		X
18		X
19	x	
20		X
21		X
22	x	
23		X
24		X
25		X
26		X
27		X
28		X
29		X
30		X
31		X
32		X

Conclusion	FAIL
-------------------	-------------


Operating requirements for surgical masks based on EN 14683: 2019+AC: 2019 standard

TEST	TYPE I	TYPE II	TYPE IIR
Bacterial filtration efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98
Differential pressure (Pa/cm ²)	< 40	< 40	< 60
Splash resistance pressure (kPa)	Not required	Not required	≥ 16
Microbial cleanliness (CFU/g)	≤ 30	≤ 30	≤ 30

Cosmetics &
Personal Care

LAB N° 1827 L

Page: 1 of 2

TEST REPORT	Refer to Analytical Report Number																				
SPONSOR	Eurofins Textile & Footwear Testing Spain																				
	C/Germán Bernácer 4																				
	03203 Elche (Alicante)																				
	SPAIN																				
TEST METHOD	Bacterial Filtration Efficiency (BFE) – EN 14683:2019+AC:2019 App B																				
TEST ITEM - INFORMATION FROM THE SPONSOR																					
PRODUCT NAME	560-2020-00007819 - Masks																				
MATRIX OF THE PRODUCT	Face Mask																				
BATCH	EUAA70-00008926	CODE	Not provided																		
EUROFINS COSMETICS & PERSONAL CARE ITALY IDENTIFICATION																					
MATERIAL ITEM ALIQUOT	N720AA0439-1																				
PARCEL REGISTRATION N.	IP-N7-2020301-AAF	RECEIVING DATE	27 Oct 2020																		
ANALYSIS STARTING DATE	28 Oct 2020	ANALYSIS ENDING DATE	29 Oct 2020																		
EXPERIMENTAL CONDITIONS	Dimension of the test specimen: 175 mm x 100 mm Size of the area tested: 49 cm ² Flow rate during testing: 28,3 l/min Inner side of the mask to the aerosol challenge.																				
PHOTO OF THE TEST ITEM																					
RESULTS	<table border="1"> <thead> <tr> <th></th> <th>RESULT</th> <th>UNIT</th> </tr> </thead> <tbody> <tr> <td>ALIQUOT 1</td> <td>83,54</td> <td>%</td> </tr> <tr> <td>ALIQUOT 2</td> <td>83,57</td> <td>%</td> </tr> <tr> <td>ALIQUOT 3</td> <td>82,61</td> <td>%</td> </tr> <tr> <td>ALIQUOT 4</td> <td>83,54</td> <td>%</td> </tr> <tr> <td>ALIQUOT 5</td> <td>84,33</td> <td>%</td> </tr> </tbody> </table>				RESULT	UNIT	ALIQUOT 1	83,54	%	ALIQUOT 2	83,57	%	ALIQUOT 3	82,61	%	ALIQUOT 4	83,54	%	ALIQUOT 5	84,33	%
	RESULT	UNIT																			
ALIQUOT 1	83,54	%																			
ALIQUOT 2	83,57	%																			
ALIQUOT 3	82,61	%																			
ALIQUOT 4	83,54	%																			
ALIQUOT 5	84,33	%																			
DETAILED RESULTS	See Addendum N. 1 (1 page)																				

*This test report may not be reproduced in part unless expressly approved in writing by Eurofins Cosmetics & Personal Care Italy Srl.
 The test results relate only to the tested items. Sampling, except specific indication on test report, is always intended to be made by the Sponsor.
 Information on the test item provided by the Sponsor are under Sponsor responsibility.*

Eurofins Cosmetics & Personal Care Italy Srl – via B.Buozzi 2, Vimodrone (Milano), Italy –
 P.IVA / VAT Number: 05533561006
 Tel: +39-022507151 – Fax: +39-0225071599 – E-mail: : InfoCosme@eurofins.com



Cosmetics &
Personal Care



LAB N° 1827 L

Page: 2 of 2

Addendum N.1

Started on: 28/10/2020

Batch: N720AA0439

Sample description: 560-2020-00007819 - Masks

Lot Number: EUAA70-00008926

Negative Control Plate Counts

	Stage 1	Stage 2	Stage 3*	Stage 4*	Stage 5*	Stage 6*	Mean
Negative Control (CFU)	0	0	0	0	0	0	0

*number of colonies adjusted with positive-hole correction table

Positive Controls Plate Counts

	Stage 1	Stage 2	Stage 3*	Stage 4*	Stage 5*	Stage 6*	Total CFU
Size of particle (µm)	7,00	4,70	3,30	2,10	1,10	0,65	
Positive Control N.1 (CFU)	203	342	1010	715	362	262	2894
Positive Control N.2 (CFU)	236	386	902	597	459	345	2925

*number of colonies adjusted with positive-hole correction table

Mean of the total plate counts of the two positive controls (CFU): 2910

Mean Particle Size (MPS)

	MPS
Positive Control N.1 (µm)	2,91
Positive Control N.2 (µm)	2,88
Mean (µm)	2,90

Test specimens Plate Counts

	Stage 1	Stage 2	Stage 3*	Stage 4*	Stage 5*	Stage 6*	Total CFU
N720AA0439-1 - Aliquot 1	4	8	41	65	153	208	479
N720AA0439-1 - Aliquot 2	3	6	31	67	163	208	478
N720AA0439-1 - Aliquot 3	4	12	30	87	172	201	506
N720AA0439-1 - Aliquot 4	6	9	38	63	154	209	479
N720AA0439-1 - Aliquot 5	4	10	41	60	153	188	456

*number of colonies adjusted with positive-hole correction table

Test specimens Bacterial Filtration Efficiency (BFE)

	BFE (%)
N720AA0439-1 - Aliquot 1	83,54
N720AA0439-1 - Aliquot 2	83,57
N720AA0439-1 - Aliquot 3	82,61
N720AA0439-1 - Aliquot 4	83,54
N720AA0439-1 - Aliquot 5	84,33

Eurofins Cosmetics & Personal Care Italy Srl – via B.Buozzi 2, Vimodrone (Milano), Italy –

P.IVA / VAT Number: 05533561006

Tel: +39-022507151 – Fax: +39-0225071599 – E-mail: : InfoCosme@eurofins.com